


# Technical Compliance Statement

## CE EMC Test Report

For the following information

Ref. File No.: C1M1106053

Product Name : Intelligent LLLT Dose Unit  
Model Number : Klas-DX series  
Brand :   
Applicant : Konftec Corporation  
Manufacturer : Konftec Corporation  
Standards :  
EN 60601-1-2:2007: Electromagnetic Compatibility  
Emission: CISPR 11:2003 +A1:2004 +A2:2006, Class B,  
EN 61000-3-2:2006 +A1:2009 +A2:2009 & EN 61000-3-3:2008  
Immunity: IEC 61000-4-2:2008, IEC 61000-4-3:2008,  
IEC 61000-4-4:2004 +Corr.1:2006 +Corr.2:2007,  
IEC 61000-4-5:2005, IEC 61000-4-6:2008, IEC 61000-4-8:2009,  
IEC 61000-4-11:2004

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council medical directive 93/42/EEC. The test data and results are issued on the EMC test report no. EM-E1000363.

Signature



Leon Liu/Deputy General Manager

Date: Sep. 08, 2011

Test Laboratory:  
AUDIX Technology Corporation, EMC Department  
TAF Accreditation No.: 1724  
Web Site: [www.audixtech.com](http://www.audixtech.com)



The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.