



## VERIFICATION OF COMPLIANCE

*This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.*

**Technical Standard: EMC DIRECTIVE 93/42/EEC  
(EN 60601-1-2)**

*(Operation Environment: Medical Electrical Equipment)*

### General Information

Applicant: KONFTEC CORPORATION  
13F-7, No.77, Sintai 5th Rd., Sec. 1, Sijhih City,  
Taipei County 221, Taiwan

### Product Description

EUT Description: Klas Portable Laser Unit  
Trade Name: Konf  
Model Number: KLAS-A; KLAS-S; KLAS-N; KLAS-E; KLAS-D



### Measurement Standard


EN 60601-1-2: 2001 + A1: 2006  
EN 61000-3-2: 2006  
EN 61000-3-3: 1995 + A1: 2001 + A2: 2005  
EN 55011: 2007 (Group 1, Class B)  
(IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000; IEC 61000-4-3: 2006; IEC 61000-4-4: 2004;  
IEC 61000-4-5: 2005; IEC 61000-4-6: 2007; IEC 61000-4-8: 1993 + A1: 2000; IEC 61000-4-11: 2004)

### Measurement Facilities

Sindian BU. Name: **Compliance Certification Services Inc.**  
No.163-1, Jhongsheng Rd., Sindian City, Taipei County 23151, Taiwan  
Tel: +886-2-22170894 / Fax: +886-2-22171029

*This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: 80312206-E*

  
  
Vince Chiang / Assistant Manager  
Date: March 18, 2008



Compliance Certification Services Inc.