



DECLARATION OF CONFORMITY

PIL No.: PIL-304151

Date: Apr. 15, 2013

PIL Title: Declaration of Conformity

Subject: Product - Low Level Laser Therapy System

Manufacturer: Konftec Corporation

Address: 12F-7, No. 77, Sec. 1, Sintai 5th Rd., Sijhih District, New Taipei City, Taiwan

Product: Low Level Laser Therapy System

Trade Name: Konf

Model Numbers: Aculas-AM-X (X=100; 102; 151; 200; 201; 273)

Aculas-HB-Y (Y=150; 250; 251; 750; 850; 950)

emLas-S (S= 150; 250; 320; 520; 650; 800; 900)

Klas-Z (Z= A; D; E; N; S; DX)

Megalas-AM-800

Megalas-HR-400

We herewith declare that the above mentioned products meet the provisions of the following EC Council Directive and Standards. All supporting documentations are retained under the premise of manufacturer.

- Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC concerning medical device (MDD 93/42/EEC)
- EN 60601-1-2: 2001 or 2007(EMC testing of Medical Electric Equipment)
- EN 55011: 1998 + A1:1999 + A2:2002
- IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000; IEC 61000-4-3: 2002 + A1: 2002; IEC 61000-4-3: 2006; IEC 61000-4-4: 2004; IEC 61000-4-5: 2006; IEC 61000-4-6: 2007; IEC 61000-4-8: 1993; IEC 61000-4-11: 2004
- EN 61000-3-2: 2006
- EN 61000-3-3: 1995 + A1: 2001

Eric Lin, EE, PhD
General Manager
Konftec Corporation

Konftec Corporation

12F-7, No. 77, Sec. 1, Sintai 5th Rd.
Sijhih District, New Taipei City 221
Taiwan



www.konftec.com

TEL: 886-2-8698-2148

FAX: 886-2-8698-2149