

Klas-N **Nasalis LLLT Unit**



User's Guide



English

Getting started/ Using Klas	2
Specifications	3
Tips and troubleshooting	4
Safety and cleaning	5

Getting started

Install your Klas with the included battery and begins to get started low level laser therapy (LLLT).

Using Klas

Klas can be used for treatment of diseases and conditions, follow the recommended dosage and treatment time to get the best efficacy. You can also find new information about using Klas on the web www.konftec.com.

Using the battery

- ☑ Use the battery that came with your Klas or any qualified standard battery or rechargeable battery.
- ☑ If you do not use your Klas for a period of time, be sure to take battery out from your Klas.

Specifications

Klas-N50

Nasalis LLLT Unit

Wavelength---660nm

Output Power---50mW (without fiber rod)
< 5mW (with fiber rod; end dispersion)

Battery-----AAA x 3

Enclosure---IP54 (NEMA 4)




About operating and storage temperatures

- ☑ Operate your Klas in a place where the temperature is always between 0 and 50° C (32 to 122° F).
- ☑ Store your Klas in a place where the temperature is always between -20 and 85° C (-4 to 185° F).

Troubleshooting

My Klas won't turn on

- ☑ Make sure Klas's power switch is on.
- ☑ If that doesn't work, check the position of battery in Klas.
- ☑ If that doesn't work, loosen the bottom cap and fasten again.
- ☑ If that doesn't work, contact your local sales office.



Learning more, service and support

There is a wealth of information about using Klas on the Web. For the latest information on Klas service and support information go to www.konftec.com

Obtaining warranty service

If the product appears to be damaged or does not function properly, please follow the advice in this booklet, and the online resources. If the unit still does not function, go to www.konftec.com for instructions on how to obtain warranty service.

Safety

Important safety instructions

Follow all instructions and warnings dealing with your Klas.

Warning

- ☑ Avoid direct eye exposure to the laser beam.
- ☑ Operation of this device, must always be supervised by an adult.



Cleaning

Follow these general rules when cleaning the outside of your Klas and its components:

- ☑ Use a damp, soft, lint-free cloth.
- ☑ Use water or alcohol to clean fiber rod.
- ☑ Don't use aerosol sprays, solvents or abrasives.

Avoid direct sunlight

Protect your Klas from long time direct sunlight.



Do not make repairs yourself

No user-serviceable parts are inside.



About handling

Your Klas may be damaged by improper storage or handling. Be careful not to drop your Klas when treating or transporting the device.



Konftec Corporation

12F-7, No.77, Sintai 5th Rd., Sec. 1,
Sijhih District, New Taipei City 221
Taiwan

www.konftec.com

© 2013 Konftec Corporation. All rights reserved.

Konf, the Konf logo and Klas
are trademarks of Konftec Corporation,
registered in Taiwan and
other countries.

UKN_30401E

