

**Asia Safety Link Inc.**

9F-1, No. 80, Sec. 2, Guang Fu Rd., San Chung City, Taipei Hsien, Taiwan  
Tel:+886-2-85123188 Fax:+886-2-29959169

# Certificate of Compliance


Certificate Number: 95-0856

Applicant : Konftec Corporation  
13F-7, No. 77, Sec. 1, Sintai 5<sup>th</sup> Rd., Sijhih City, Taipei County 221

Product : Multi-Channel Low Level Laser Therapy System

Model/Type : Aculas-AM-x, where x = 100, 151, 102, 200, 201 or 273  
Aculas-HB-y, where y = 150, 250, 251, 750, 850 or 950

Electrical Rating: 9Vdc, 4A

Trade Mark:  **Konf**

Standards applied: IEC 60601-1:1988 + A1:1991 + A2:1995

The tested samples of the above products are in conformity with the technical provisions of the following

Date Issued: September 25, 2006

Approve & Authorized Signer: \_\_\_\_\_



Jeff Chang



## VERIFICATION OF COMPLIANCE

*This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.*

**Technical Standard: EMC DIRECTIVE 93/42/EEC  
(EN 60601-1-2)**

*(Operation Environment: Medical electrical Equipment)*

### **General Information**

Applicant: KONFTEC CORPORATION  
13F-7, No. 77, Sintai 5th Rd., Sec. 1, Sijhih City,  
Taipei County 221, Taiwan

### **Product Description**

EUT Description: Multi-channel Low Level Laser Therapy System  
Trade Name: Konf  
Model Number: Aculas-AM-X (X=100; 151; 102; 200; 201; 273);  
Aculas-HB-Y (Y=150; 250; 251; 750; 850; 950)


### **Measurement Standard**

EN 60601-1-2: 2001  
EN 55011: 1998 + A1: 1999 + A2: 2002  
(IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000; IEC 61000-4-3: 2002 + A1: 2002;  
IEC 61000-4-4: 1995 + A1: 2000 + A2: 2001; IEC 61000-4-5: 1995 + A1: 2000;  
IEC 61000-4-6: 1996 + A1: 2000; IEC 61000-4-8: 1993 + A1: 2000; IEC 61000-4-11: 1994 + A1: 2000)  
EN 61000-3-2: 2000  
EN 61000-3-3: 1995 + A1: 2001

### **Measurement Facilities**

Laboratory Name: **Compliance Certification Services Inc.**  
No. 165, Chung Sheng Road, Hsin Tien City, Taipei, Taiwan, R.O.C.  
Tel: +886-2-22170894 / Fax: +886-2-22171029

*This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: 60615202-E*

  
Vince Chiang / Assistant Manager  
Date: August 9, 2006