



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.

**Technical Standard: EMC DIRECTIVE 93/42/EEC
(EN 60601-1-2)**

(Operation Environment: Medical Electrical Equipment)

General Information

Applicant: KONFTEC CORPORATION
13F-7, No.77, Sintai 5th Rd., Sec. 1, Sijhih City,
Taipei County 221, Taiwan

Product Description

EUT Description: Klas Portable Laser Unit
Trade Name: Konf
Model Number: KLAS-A; KLAS-S; KLAS-N; KLAS-E; KLAS-D

Measurement Standard

EN 60601-1-2: 2001 + A1: 2006

EN 61000-3-2: 2006

EN 61000-3-3: 1995 + A1: 2001 + A2: 2005

EN 55011: 2007 (Group 1, Class B)

(IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000; IEC 61000-4-3: 2006; IEC 61000-4-4: 2004;

IEC 61000-4-5: 2005; IEC 61000-4-6: 2007; IEC 61000-4-8: 1993 + A1: 2000; IEC 61000-4-11: 2004)


Measurement Facilities

Sindian BU. Name: Compliance Certification Services Inc.
No.163-1, Jhongsheng Rd., Sindian City, Taipei County 23151, Taiwan
Tel: +886-2-22170894 / Fax: +886-2-22171029

This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: 80312206-E

Vince Chiang / Assistant Manager

Date: March 18, 2008



Compliance Certification Services Inc.